


PRODUCT DATA SHEET

Revision 1/	Composite Serological Pipette	
Validfor Item-No.:	1040311/1041311/1042311/1043311/1043301/1044301/1045301/1040310/1041310/1042310/1043310/1043300/1044300/1045300	

1.	Description / Specification	
1.1	Description	Composite serological pipette PS, clear 1040311 Composite serological pipette, 1ml, scale 1/100, stretch type, yellow color ring 1041311 Composite serological pipette, 2ml, scale 1/50, stretch type, green color ring 1042311 Composite serological pipette, 5ml, scale 1/10, stretch type, blue color ring 1043311 Composite serological pipette, 10ml, scale 1/10, stretch type, orange color ring 1043301 Composite serological pipette, 10ml, scale 1/10, welding type, orange color ring 1044301 Composite serological pipette, 25ml, scale 2/10, welded type, red color ring 1045301 Composite serological pipette, 50ml, scale 1/2, welding type, purple color ring 1040310 Composite serological pipette, 1ml, scale 1/100, stretch type, yellow color ring 1041310 Composite serological pipette, 2ml, scale 1/50, stretch type, green color ring 1042310 Composite serological pipette, 5ml, scale 1/10, stretch type, blue color ring 1043310 Composite serological pipette, 10ml, scale 1/10, stretch type, orange color ring 1043300 Composite serological pipette, 10ml, scale 1/10, welded type, orange color ring 1044300 Composite serological pipette, 25ml, scale 2/10, welding type, red color ring 1045300 Composite serological pipette, 50ml, scale 1/2, welding type, purple color ring
1.2	Dimensions	Flask: see Customer Drawing 1040311-1045301: bagged 1040310-1045300: boxed
1.3	Volume	See 1.1Description
1.4	Material / Resin	Serological pipette: PS (Polystyrene) Filter: PE (Polyethylene)
1.5	Colour	Serological pipette: clear Filter: white
1.6	Sterilization	SAL 10 ⁻⁶
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for pipetting of liquids to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only Expiry date and Lot-No. printed on the outer box

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	For application: 0°C to +40 °C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Shelf life	3 years
2.6	Other Information	-

3.	Packaging	1040311	1041311	1042311/ 1043311/ 1043301	1044301	1045301	1040310/ 1041310	1042310/ 1043310/ 1043300	1044300/ 1045300
3.1	Pieces / Bag	50	50	50	25	25	200	100	50
3.2	Pieces / Box	20	10	4	8	4	4	4	4
3.3	Lot-No.	DDD YY XXX (day, year, product No.)							
3.4	Other Information	Certificate of Quality to download							

4.	Other Information	
4.1	Laboratory use only	

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Zhejiang Saining Biotechnology Co., Ltd. This document may not be reproduced for any reason without written permission from Zhejiang Saining Biotechnology Co., Ltd. All rights of design, invention, and copyright are reserved.
Revision 1	Date 1 June 2024	Date 1 June 2024	Date 1 June 2024	
Date 01.06.2024	Name Chen Anjing	Name Zhang Heng	Name Wang Kedong	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.